

Interactive comment on “Moving beyond the age-depth model paradigm in deep sea palaeoclimate archives: dual radiocarbon and stable isotope analysis on single foraminifera” by Bryan C. Lougheed et al.

Bryan C. Lougheed et al.

bryan.lougheed@geo.uu.se

Received and published: 23 November 2017

Dear Julia Gottschalk,

Thank you for interest in our manuscript, your encouraging words and for your contribution to the discussion forum of this manuscript.

Indeed, when analysing for the rare radioisotope ^{14}C on such small samples, one must be vigilant for any possibility of modern ^{14}C contamination, as pointed out in the publications you have cited in your short comment: Brown and Southon (1997); Hua

C1

et al. (2004). Those publications specifically refer to possible contamination during the graphitisation process, which we did not use. The method which we use involves acidifying sample carbonate to sample CO_2 in He-flushed, sealed septa vials for direct measurement in the MICADAS AMS (e.g. Wacker et al., 2013). This method has much smaller sample mass requirements than the graphitisation process, meaning that it is possible to measure much smaller sample masses on the AMS. However, we agree that the general principle of small sample masses being relatively more susceptible to contamination of course applies, just that the mass threshold is much lower than in the case of graphitisation. We measured IAEA-C1 standard powder, as well as single Eemian (i.e. assumed ^{14}C “blank”) foraminifera from the same species and sediment core. The masses of these blank foraminifera samples (min: $12.8 \mu\text{g C}$, max: $24.0 \mu\text{g C}$, mean: $19.0 \mu\text{g C}$) were of similar range to that of the foraminifera used for the study (min: $8.4 \mu\text{g C}$, max: $47.2 \mu\text{g C}$, mean: $19.8 \mu\text{g C}$). Procedural blank foraminifera values ranged between $\sim 31,000$ and $\sim 41,000$ ^{14}C yr BP and were generally not as good as those for IAEA-C1 material. There was no correlation between sample mass and ^{14}C value, neither in the case of the IAEA-C1 material nor the blank forams.

The specific blank correction value used was $F^{14}\text{C} = 0.0095 \pm 0.00203$ (~ 37500 ^{14}C yr BP). This blank value, along with its uncertainty, was applied using the BATS software (Wacker et al., 2010). For a smaller, second run of samples (ETH-74XX) it was possible to use a much lower blank ($F^{14}\text{C} = 0.0033 \pm 0.00101$).

You are correct that the variability of process blanks and the reduced measurement sensitivity reduce, for now, for now, the ability to precisely measure older samples. Moreover, as you suggest, the variation also reduces the precision somewhat on younger dates (e.g., in the $<20,000$ ^{14}C yr BP range). Compared to a constant blank value of $37,500$ ^{14}C yr, variable blank values between $\sim 31,000$ and $\sim 41,000$ ^{14}C yr would affect ^{14}C ages of Holocene samples (i.e. the bulk of our samples) by between $\sim 0.9\%$ and $\sim 3.5\%$, thus increasing uncertainty. Due to the relatively poorer counting statistics inherent in ^{14}C dating single forams, our reported ^{14}C dates already have

C2

quite a low precision (between 1% and 4%) anyway. Moreover, the uncertainties associated with ^{14}C reservoir age reduce chronological precision by a further 7.5% on average. Therefore, the variation in blank value is not significant in determining the quantified PDSM in T86-10P (which is on the order of thousands of years). In general, dating single forams sacrifices precision to gain accuracy. Such an approach is especially useful in sediment cores with high PDSM and/or very low SAR, where large multi-specimen samples would yield a good ^{14}C signal with low error, but all information regarding the large intra-sample ^{14}C age heterogeneity would be lost.

Pre-treatment options for single forams is something that we are continually looking into. It is our intention that in future, a routine and (automated) prep system for single foraminifera will be developed, with the ability to carry out pretreatment online and split gaseously into stable isotope and ^{14}C fractions. However, this depends on access to project funding.

Thank you also for pointing out that no information is provided regarding sample size (i.e. mass) in the supplemental table. We agree that this information should definitely be provided in the case of our study, as we detail new methods. We will include information regarding sample $\mu\text{g C}$ as reported by the MICADAS AMS in the supplemental table. You might note that some of the larger sample masses with ^{14}C data have no corresponding stable isotope data. In principle it should have been possible to also report stable isotope data for these larger samples, but we suffered a failed IRMS run at another laboratory (not at our affiliation institutions).

Once again, thank you for your interest in our manuscript and specifically your helpful comments requesting more information about the blank correction process. We will better highlight this process (as explained above) in the updated version of the manuscript, which will certainly serve to significantly improve the reader experience.

On behalf of the co-authors, kind regards,

Bryan Lougheed

C3

Wacker, L., Christl, M. and Synal, H.A., 2010. BATS: a new tool for AMS data reduction. *Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms*, 268(7), pp.976-979.

Wacker, L., Lippold, J., Molnár, M. and Schulz, H., 2013. Towards radiocarbon dating of single foraminifera with a gas ion source. *Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms*, 294, pp.307-310.

Interactive comment on *Clim. Past Discuss.*, <https://doi.org/10.5194/cp-2017-119>, 2017.

C4